

LIMITED TENDER # 06/2015

Sealed Quotation/Tenders quoting rates for the supply of stores as detailed below to National Institute for Biotechnology and Genetic Engineering (NIBGE), FAISALABAD, are invited as under:-

Limited Tender #	Item	Reference File No.	Date & Time of Tender Opening
06/2015	Centrifuge Bench Top etc	Proc-5(275)/2015 /NIBGE	04-06-2015 at 1430 hours

The following terms and conditions have been laid down for suppliers for submitting their quotations to NIBGE, Faisalabad.

- (I) The quotations should be submitted showing description/ specifications/ drawings along with signatures of dealer/ suppliers duly affixed with his seal. The suppliers are not allowed to change or modify the given specifications.
- (II) Additional information & Literature may also be attached and position of availability of stores in stock, indicated before signing the quotations. Samples may be attached, if required.
- (III) Quotation/Tender should be sent in sealed cover duly marked with **File No. and opening dated**. Quotations received without the marking on the Envelope will not be considered.
- (IV) Quotation should normally be valid for a period of 90-days from the date of opening of quotations and it will be presumed that the Firms/dealers will supply the items on credit.
- (V) The bill will be paid at the earliest possible, provided the supplied store is of the ordered's specifications, in serviceable condition & correct quality.
- (VI) All amendment/over writing of rates in quotation may please be attested failing which the quotations will not be considered.
- (VII) The rates are to be quoted on F.O.R. NIBGE, FAISALABAD, basis.
- (VIII) In any condition contrary to these terms detailed above are laid down by the supplier, such condition would not be acceptable or binding on NIBGE.
- (IX) **Earnest Money @ 2% of the total quoted value (including 17% GST) in the shape of Deposit-at-Call duly affixed Revenue Stamps according to its value (Cheque /Pay Order/ Bank Draft/ Demand Draft not acceptable) in favour of Accounts Officer, NIBGE, FAISALABAD must accompany the quotation. The bids not supported with CDR OR with less amount earnest money shall be rejected at spot.**
- (X) The quotations may be sent through POST/COURIER SERVICE to the undersigned OR by hand at R & I (NIBGE) Faisalabad before due date. Quotations shall be received at **1400** hours and opened at **1430** hours.
- (XI) The Competent Authority reserves the right to accept/reject any of the quotations or part thereof without assigning any reason.
- (XII) The quotations of those firms will be accepted which are registered with Sales Tax Department having Sales Tax Registration which should clearly be mentioned on quotation.
- (XIII) The rates should be in full rupee inclusive of 17% General Sales Tax. If GST not mentioned as **Exclusive or Exempted** in quotation, the quoted rates shall be considered as **Inclusive** of 17% GST. Any decimal figure shown with quoted rates shall be considered as deleted/ignored.

NOTE: Writing as "GST as per Govt. rules" will also be deemed as misleading and shall be considered as inclusive of GST.

- **All the firms who are interested to participate must be registered with any academic or research institute working under Government of Punjab and should not be black listed by any establishment of Pakistan Atomic Energy Commission of Pakistan.**

Admin Officer (Procurement)
NIBGE, P.O.Box # 577, Jhang Road, Faisalabad.
Ph: 0092-41 9201316-20 & 22/Ext. 201,206/(0092-41) 9201472

Sr.	Description	Qty
1	<p><u>Centrifuge bench top:</u> Specification and features Max speed 15000rpm (300- 15000rpm), 10 rpm increments. Max. RCF: 21380xg, 10xg increments. Run Time: 30secs to 99 mins. Timer or Continuous Operation. Voltage: Single-phase, 60Hz, 5A Dimensions: Should not be greater that (W) 11.0 x (1,) 14.3 x (H) 10.4 inches. Large LCD display provides all visual information. Should be supplied with a 24 place 1.5/2.0 ml bio-safe aluminum rotor. High quality aluminum alloy Bio-containment rotors, safe and autoclavable. Whisper quiet operation. Brushless motor drive, maintenance free, Double cooling air design, minimizes rotor heating, Centrifuge should be compatible with rotors of Capacity 0.5ml and 0.2ml tubes and 4 x 8 tube PCR strip rotor. Pulse operation for quick spins, automatic door lock And over speed detection. Origin of supplier company must be US or Europe Or equivelant.</p>	1